

In Reference To: O-1319
Case No.: 45671 SDG(s): F6R18

**Contract Laboratory Program
REGIONAL/LABORATORY COMMUNICATION SYSTEM**

Resubmission Request

| | |
|--------------------------|----------------------|
| Laboratory Name: | EQI |
| Lab Contact: | Robert Zhu |
| Region: | 6 |
| Regional Contact: | Raymond Flores - EPA |
| ESAT Reviewer: | Wallace Doong - ESAT |

In reference to data for the fraction(s) :

CSF Deliverables LMVOA SVOA/SVOA-SIM

Summary of Questions/Issues:

CSF Deliverables

The "Date" for "Completed By" was omitted on Form DC-2-7 (page 15). Please correct and resubmit this page.

LMVOA

The "Level" should be "MED" for Form 6A (pages 270, 271, 301, and 302), Form 7A (pages 369, 370, 377, 378, 385, and 386), and Form 8A (pages 47 and 48). Please correct and resubmit these pages with proper pagination.

SVOA and SVOA--SIM

1. **SVOA-SIM:** For Form 8A (pages 960 to 965), the IS3 should be phenanthrene-d10 and results for IS6 (1,4-Dichlorobenzene-d4) were omitted for all samples. Please correct and resubmit these pages with proper pagination.
2. **SVOA and SVOA-SIM:** The recoveries for 1,4-Dioxane-d8 (DMC1) were outside the QC limits for SVOA method blank SBLK12 and SVOA-SIM method blanks SBLK13 and SBLK38, but contract required reanalyses and/or re-extractions (SOM02.3, p. D-54/SVOA, sec. 12.1.2.5.2 and D-55/SVOA, sec. 12.1.2.6.3) were not performed. Please submit additional information or explain.

NOTE: Any laboratory resubmission should be submitted either as an addendum to the original CSF with a revised Form DC-2 or submitted as a new CSF with a new Form DC-2 except for replacement pages (SOM02.3, p. B-11, sec. 2.2.2). Custody seals are required only for regular mail shipments.

Please respond to the above items **within 5 business days** (SOM02.3, p. B-11, sec. 2.2.1) by e-mail to Flores.Raymond@epa.gov. If you have any questions, please contact Mr. Flores at 281-983-2139.

Distribution: (1) Lab Copy, (2) Region Copy, and (3) ESAT Copy